



OFFICE OF THE MISSION DIRECTOR
NATIONAL HEALTH MISSION, ASSAM
SAIKIA COMMERCIAL COMPLEX, CHRISTAN BASTI, G.S ROAD, GUWAHATI-781005
PH.NO : 0361-2363062 ; TELE FAX : 0361-2363058
Website: www.nrhmassam.in e-mail: misnrhm.assam@gmail.com

NO: NHM-18019/1/2019-PROC-NHM/99-P

DATE: 2/7/2019

CORRIGENDUM NO.2
NOTICE INVITING TENDER
(NATIONAL COMPETITIVE BIDDING)

This is with reference to the TENDER NO: NHM/18019/1/2019-PROC-NHM/6153, DATE: 18/6/2019 & Corrigendum No. NHM/18019/1/2019-PROC-NHM/ 6729, DATE: 21/6/19

The following amendments in the tender should be taken note of prior to submission of bids and any deviation will not be accepted.

- 1.** The specification of the Dengue NS 1 Antigen Detection Kit by Elisa may be read as

Sl No.	Name of item	Specification	Packing Size
1	Dengue NS 1 Antigen Detection Kit by Elisa	<ol style="list-style-type: none">1) The Elisa Kit should be designed for qualitative detection on Dengue NS1 antigen of all 4 Dengue serotypes in human serum.2) The kit should be provided with the following materials and reagents:<ol style="list-style-type: none">a) Anti -NS1 Antibody Coated Breakway Microwells (12 X 8 =96). Desiccant should be provided for storing the unused microwells which are to be resealed immediately.b) Horseradish peroxidise conjugated Anti-NS1 monoclonal antibody with preservative.c) Chromogenic substrate in buffer.d) Positive control in the form of recombinant antigen with preservatives and antibodies.e) Negative control in the form of confirmed negative human serum with preservatives and antibodies.f) Calibrators in the form of recombinant antigen with preservatives and antibiotics (optional)	1 Unit = A single kit of 1 X 96 test

		<p>g) Sample diluents. h) Wash Buffer.</p> <p>3) The time required for performing the test for detection of Dengue NS1 antigen by Elisa should range between 2- 4 hours.</p> <p>4) The Elisa kit for detection of Dengue NS1 should have a sensitivity of > 90% and a specificity of >95% taking RT-PCR as the gold standard.</p> <p>5) The kit should have a self - life of at least 6 months when stored at an ambient temperature of 2⁰ C -8⁰ C Transportation should be under cold chain.</p> <p>6) To Provide stopping solution along with the kit.</p>	
--	--	--	--

The last date of Bid submission will be extended till 11/07/2019 up to 2.00 pm.
All other terms & Conditions of the above referred tender shall remain unchanged.

Sd/-
Mission Director, NHM, Assam