

OFFICE OF THE MISSION DIRECTOR
NATIONAL HEALTH MISSION, ASSAM

Saikia Commercial Complex, Sri Nagar Path, Near Post Office Bus Stand, G. S. Road, Christianbasti, Guwahati - 781005
Website: <https://nhm.assam.gov.in> :: Phone No: 0361-2340236/39 :: Email Id: mdrnhmasm@gmail.com

No. NHM-18018/3/2021-Procurement-NHM-National Health Mission/ECF:174064/ 99- P DATE: 28/9/21

CORRIGENDUM NO.1

e-TENDER FOR RATE CONTRACT OF Salt Testing Kit

This is with reference to NHM-18018/3/2021-Procurement-NHM-National Health Mission/
ECF:174064/ 18603, Dated 1-9-21.


The following amendments in the tender may be taken note of prior to submission of bids.

1. Bidder should submit 6-10 nos. of sample Kit along with the Technical Bid. Sample submission is mandate & without the same the tender will be consider as rejected. Financial Bid will be opened of the technically accepted plus sample cleared bidder only for Salt Testing Kit.
1. In addition to the certification already mentioned in the original Bid document i.e ISO 9001: 2015/ISO 9001: 2008/ ISO 13485/QMS (Quality Management Certificate)/ GMP/WHO-GMP, now the test kit of the manufacturer must be evaluated & validated by WHO/UNICEF/MI and /or National Level Laboratories such as National Institute of Nutrition-Hyderabad, National Centre for Disease Control-Delhi, All India Institute of Medical Science-New Delhi, All India Institute of Hygiene & Public Health-Kolkata, Central Food Technological Research Institute-Mysore, Indian Council of Medical Research & Council of Scientific & Industrial Research Laboratories.
The validation should include tests for the Quality, packaging, ready to use testing (drop by drop), stability at various places, shelf life under sealed condition as well as open condition. The testing laboratory should submit a detailed report about all the test parameters including how they vary under different field condition.
The same should be supported by documentary evidence/certificate from the Competent Authority.
2. Shelf life: The shelf life of the salt testing Kit should be at least 1 year & when the vial is opened it should not be less than 4-6 months, supported by documentary evidence/certificate from the Competent Authority.
3. Pack Size: Each salt testing Kit should be independently packed, not more than 10 Kits in a secondary packing, supported by documentary evidence/certificate from the Competent Authority.
4. Manufacturers of the Salt Testing Kit should have an experience of 5 years of manufacturing & marketing, supported by documentary evidence/certificate from the Competent Authority.

The start date of submission of Tender will be extended as below:

1. Start date: will be extended from 29/09/2021 (2.00 pm) to 4/10/2021 (2.00 pm)
2. End date: will be extended from 6/10/2021 (2.00 pm) to 9/10/2021 (2.00 pm)

All other terms & conditions of the tender under reference will remain unchanged.


28/9/21
Mission Director, NHM, Assam


Memo No. NHM-18018/3/2021-Procurement-NHM-National Health Mission/ECF:174064/

98-P

DATE: 28/9/21

Copies to the:

1. Principal Secretary to the Govt. of Assam, Health & F.W. Deptt. for kind information.
2. Director of Health Services, Hengerabari, Guwahati -for information.
3. Director Medical Education, Six Mile, Khanapara, Guwahati -for information.
4. SPO, NIDDCP, Assam-for information.
5. MIS Manager, NHM, with a request to upload the Corrigendum Notice in the NHM Website.
6. SME, NHM, with a request to publish the Corrigendum Notice in leading national & local newspapers.
7. P.S to the Honb'le Minister, H &FW, Govt. of Assam for kind appraisal of Honb'le Minister.


28/9/21

Mission Director, NHM, Assam