



OFFICE OF THE MISSION DIRECTOR
National Health Mission, Assam
Saikia Commercial Complex, Srinagar Path, Christian Basti,
G S Road, Guwahati-781005, Assam

Website : www.nrhmassam.in

e-mail : misnrhm.assam@gmail.com

Ref: NHM/PROC/PHYSIO-EQUIP/2175/2014-15/6861

Date: 23/07/2014

CORRIGENDUM NO-1

NOTICE INVITING TENDER FOR PHYSIOTHERAPY EQUIPMENTS & INSTRUMENTS

(NATIONAL COMPETITIVE BIDDING)

This is with reference to tender No. NHM/PROC/PHYSIO-EQUIP/2175/2014-15/5367 Dated 03/07/2014 for supply & installation of Physiotherapy Equipments & Instruments. The following amendments in the tender may be taken note of prior to submission of bids:

1. **CLAUSE 1 (LAST DATE FOR RECEIPT OF TENDERS) IS AMENDED TO READ AS FOLLOWS**

Sealed Tenders in two separate covers {Technical bid (Cover "A") and Price Bid (Cover "B")} will be received up to 31/07/2014 till 2 PM by the Mission Director, National Health Mission, Saikia Commercial Complex, Srinagar path, Christianbasti, G S Road, Guwahati-781005, Assam

2. **CLAUSE 4 (B) & (C) (ELIGIBILITY CRITERIA) IS AMENDED TO READ AS FOLLOWS:**

4(B): Average Annual turnover of the bidder (who is not a manufacturer) in the last three financial years i.e. 2010-11, 2011-12, 2012-13 from sales of Physiotherapy Equipment / Hospital Furniture & Fixtures/Medical Equipments/Instruments shall not be less than **Rs 75.00 Lakhs (Rupees Seventy Five Lakhs).**

4(C): Manufacturer participating as bidder or manufacturer from whom the **tendered items would be sourced by the dealer should have an average Annual Turnover of **Rs 1 (One) Crore** in the last 3 financial years i.e. 2010-11, 2011-12, 2012-13.**

3. **ANNEXURE I (TECHNICAL CUM COMPLIANCE SPECIFICATION SHEET) IS AMENDED TO READ AS FOLLOWS:**

ITEM NO 3: (SHORT WAVE DIATHERMY UNIT): CLAUSE 2.1:

A device using electromagnetic energy in the shortwave frequency range (27.12 MHz) for therapeutic purposes. The unit includes electrodes, the shortwave generator, and all associated electronics, controls and enclosures.

ITEM NO 5: (ULTRASOUND THERAPY UNIT (SINGLE HEAD)): CLAUSE 3.1:

- Frequency of 1 MHz
- Intensity of 0-3 w/cm² with display along with timer and one water proof treatment heads, one large up to 5 cm
- Output mode continuous and pulsed
- Output power 15w in continuous mode and 21w in pulse mode
- Pulse frequency 100Hz
- Output frequency 1 MHz
- Timer 0-15 minutes, pre-settable.
- Digital display for Power intensity
- Patient safety circuit

All other terms and condition shall remain unchanged.

Sd/-
Mission Director,
National Health Mission, Assam